

Computer-Based Classification of Dermoscopy Images of Melanocytic Lesions on Acral Volar Skin

Hitoshi Iyatomi¹, Hiroshi Oka², M Emre Celebi³, Koichi Ogawa¹, Giuseppe Argenziano⁴, H Peter Soyer⁵, Hiroshi Koga⁶, Toshiaki Saida⁶, Kuniaki Ohara⁷ and Masaru Tanaka⁸

We describe a fully automated system for the classification of acral volar melanomas. We used a total of 213 acral dermoscopy images (176 nevi and 37 melanomas). Our automatic tumor area extraction algorithm successfully extracted the tumor in 199 cases (169 nevi and 30 melanomas), and we developed a diagnostic classifier using these images. Our linear classifier achieved a sensitivity (SE) of 100%, a specificity (SP) of 95.9%, and an area under the receiver operating characteristic curve (AUC) of 0.993 using a leave-one-out cross-validation strategy (81.1% SE, 92.1% SP; considering 14 unsuccessful extraction cases as false classification). In addition, we developed three pattern detectors for typical dermoscopic structures such as parallel ridge, parallel furrow, and fibrillar patterns. These also achieved good detection accuracy as indicated by their AUC values: 0.985, 0.931, and 0.890, respectively. The features used in the melanoma-nevus classifier and the parallel ridge detector have significant overlap.

Journal of Investigative Dermatology advance online publication, 6 March 2008; doi:10.1038/jid.2008.28

INTRODUCTION

In non-white populations, almost half of the melanomas are found in acral volar areas and nearly 30% of melanomas affect the sole of the foot (Saida *et al.*, 2002). Melanocytic nevi are also frequently found in their acral skin and approximately 8% of Japanese have melanocytic nevi on their soles (Saida *et al.*, 2004). The authors reported that about 90% of melanomas have the parallel ridge pattern and 70% of melanocytic nevi have the parallel furrow pattern. They concluded that the parallel ridge pattern is the most typical characteristic of acral volar melanomas.

Despite the use of dermoscopy, diagnosis is subjective and the accuracy of expert dermatologists in diagnosing melanoma is still estimated to be 75–84% (Argenziano *et al.*, 2003).

For the reasons cited above, several groups have developed automated analysis procedures to overcome these problems

and reported high levels of diagnostic accuracy (Ercal *et al.*, 1994; Seidenari *et al.*, 1998; Elbaum *et al.*, 2001; Ganster *et al.*, 2001; Rubegni *et al.*, 2002; Hoffmann *et al.*, 2003; Blum *et al.*, 2004; Burroni *et al.*, 2005; Celebi *et al.*, 2007). However, automatic detection of the parallel ridge and parallel furrow patterns is often difficult because of the wide variety of dermoscopy images, and there is no published method on computerized classification of this diagnostic category.

We developed a new diagnostic classifier for acral volar lesions and for our web-based melanoma diagnostic system (Oka *et al.*, 2004). Our method parameterizes the lesions objectively and evaluates them instead of detecting the ridges or furrows in the images directly.

RESULTS

Tumor area extraction and observed dermoscopic patterns

Our computer-based fully automated tumor area extraction method (Iyatomi *et al.*, 2006) successfully extracted tumor area in 199 out of 213 cases (Table 1). Samples of dermoscopy images are shown in Figure 1.

Out of 169 nevi, parallel ridge, parallel furrow, and fibrillar patterns were found in 5, 133, and 49 cases, respectively. A total of 11 cases of nevi had no specific patterns and 28 nevi had both parallel furrow and fibrillar patterns. One nevus had both parallel ridge and fibrillar pattern. In 30 melanomas, parallel ridge, parallel furrow, and fibrillar patterns were found in 24, 2, and 1 cases, respectively. Five melanomas had no specific patterns and one melanoma had all three patterns.

Classifier and the performance

Using principal component analysis (PCA), a total of 428 image parameters were transformed into 198 orthogonal

¹Department of Electronic Informatics, Hosei University Faculty of Engineering, Koganei, Tokyo, Japan; ²Section of Cosmetic Surgery, Shiromoto Clinic, Shinjuku, Tokyo, Japan; ³Department of Computer Science, Louisiana State University, Shreveport, Louisiana, USA; ⁴Department of Dermatology, Second University of Naples, Naples, Italy; ⁵Queensland Institute of Dermatology, University of Queensland, Brisbane, Queensland, Australia; ⁶Department of Dermatology, Shinshu University School of Medicine, Matsumoto, Nagano, Japan; ⁷Department of Dermatology, Toranomon Hospital, Minato, Tokyo, Japan and ⁸Department of Dermatology, Tokyo Women's Medical University Medical Center East, Arakawa, Tokyo, Japan

Correspondence: Dr Hitoshi Iyatomi, Department of Electronic Informatics, Faculty of Engineering, Hosei University, 3-7-2 Kajino-cho, Koganei, Tokyo 184-8584, Japan. E-mail: iyatomi@hosei.ac.jp

Abbreviations: AUC, area under the receiver operating characteristic curve; PC, principal component; PCA, principal component analysis; ROC, receiver operating characteristic; SE, sensitivity; SP, specificity

Received 20 August 2007; revised 20 November 2007; accepted 22

December 2007

principal components (PCs) without information loss. From these PCs, we selected the effective ones for each classifier. Table 2 summarizes the number of selected PCs for each classification model, the determination coefficient R^2 , the standard deviation of mean estimated error E , and the order number of the first 10 PCs lined by the selected sequence by stepwise input selection method.

In the melanoma-nevus classifier, many significant (small numbered) PCs were found in the first 10 selected features.

Table 1. Dermoscopy images used in this experiment

Lesion type	No. of images correctly extracted ¹ (in Japanese)	No. of patients correctly extracted ² (in Japanese)
Nevus	169/176 (144/151)	166/170 (141/145)
Melanoma	30/37 (29/36)	18/20 (17/19)
Total	199/213 (174/188)	184/189 (159/164)

¹The number of images that were used to correctly extract their tumor areas by computer-based tumor area extraction algorithm (Iyatomi *et al.*, 2006). Numbers in parenthesis represent those of Japanese.

²The number of patients whose images were used to correctly extract their tumor areas by extraction algorithm.




Figure 1. Samples of acral volar dermoscopy images. (a) Nevus, (b) melanoma, (c) nevus (could not be extracted by our tumor extraction algorithm), (d) melanoma (could not be extracted by our tumor extraction algorithm).

The parallel ridge and parallel furrow detectors were also composed of significant PCs. On the other hand, the fibrillar pattern detector showed a different trend.

The leave-one-out cross-validation results for these four models are shown in Table 3.

The sensitivity (SE) and specificity (SP) values shown are those that have the maximum product. The positive predictive value and negative predictive value shown are at the same diagnosis threshold. The numbers in parentheses represent the performance when 14 unsuccessful extraction cases are considered as false classification. The corresponding receiver operating characteristic (ROC) curve is shown in Figure 2.

On the other hand, the identification performance for melanoma and three typical structures using a 10-fold cross-validation was 0.991, 0.982, 0.921, and 0.883 in area under the ROC curve (AUC), respectively, and the performance by a half-and-half test was 0.991, 0.971, 0.889, and 0.839 in AUC, respectively.

Table 4 summarizes the major constituting image features (top five in magnitudes: note that each feature was normalized to $N(0, 1)$) of typical significant PCs, their linguistic expression, their contribution ratio (ratio of the corresponding eigenvalue to the total eigenvalues: $\lambda_i/\sum\lambda_i$), and their classification performance using a linear model. ($y = ax + b$;

Table 3. Identification performance of acral volar melanoma and three typical structures using leave-one-out cross-validation strategy

Classifier type	SE (%) ¹	SP (%) ¹	PPV (%) ¹	NPV (%) ¹	AUC
Melanoma	100 (81.1 ²)	95.9 (92.1 ²)	81.1	95.9	0.993
Parallel ridge pattern	93.1	97.7	87.1	98.8	0.985
Parallel furrow pattern	90.4	85.9	93.1	80.9	0.931
Fibrillar pattern	88.0	77.9	57.1	95.1	0.890

AUC, area under the receiver operating characteristic curve; NPV, negative predictive value; PPV, positive predictive value; SE, sensitivity; SP, specificity.

¹At which SE × SP is maximum value.

²When 14 unsuccessful extraction cases are treated as false classification.

Table 2. Modeling performance of developed classifiers and their constituting PCs

Classifier type	No. of inputs	R^2	E	Selected PCs (first 10) ¹
Melanoma	45	0.807	0.315	2, 9, 6, 1, 3, 15, 91, 40, 20, 98
Parallel ridge pattern	40	0.736	0.363	2, 9, 1, 6, 3, 59, 20, 88, 77, 33
Parallel furrow pattern	35	0.571	0.614	6, 2, 145, 15, 3, 98, 70, 24, 59, 179
Fibrillar pattern	24	0.434	0.654	106, 66, 56, 145, 137, 94, 111, 169, 131, 5

PC, principal component.

R^2 is the determination coefficient adjusted by degree of freedom between average of physician and the computer-based results.

E is the standard deviation of the average estimated error between the training signal and the computer-based results.

¹The first 10 selected PCs by incremental stepwise method in selected sequence. The number represents the order of PCs. The smaller values indicate the more significant component.

y is target value, x is the corresponding PC, a is the coefficient, and b is the constant bias value.)

Note that in Table 4 the image features are abbreviated. For example, “diffX_T80” represents the difference of center of mass in the x direction (diffX) between the original tumor area and the area with intensity lower than a threshold of 80 (T80). Similarly, “corr90_δ1/2” represents the texture correlation (corr) in the tumor area in the vertical direction (90 degree) as measured by a co-occurrence matrix with distance parameter δ , the half of the major axis length of the tumor object. R, G, and B indicate red, green, and blue channel values, respectively. H, S, and V represent the hue, saturation, and intensity in the HSV color space, respectively.

DISCUSSION

Automatic tumor area extraction

In 14 cases (7 nevi and 7 melanomas), our automated tumor area extraction algorithm failed. This was due to the size of

the tumor being larger than about 70% of the dermoscope field. Our algorithm is mainly for early melanomas that usually fit in the frame. As larger lesions are relatively easy to diagnose, we deem that computer-based screening is not necessary. The false extraction rate for melanomas was higher (19%) than that of nevi (4%). However, our algorithm successfully extracted 93.4% of the tumor areas.

Analyzing results for each dermoscopic pattern

Our method achieved high detection rate for melanoma and parallel ridge pattern. Almost equivalent results were obtained using leave-one-out cross-validation, 10-fold cross-validation, and half-and-half test, and therefore the results can be considered as reasonable. We think that the accurate recognition of parallel ridge pattern is important for the accurate detection of melanoma.

The melanoma classifier and the parallel ridge detector have many common PCs. In particular, the top five PCs for the two were completely the same. Note that parameters chosen early in the stepwise feature selection were thought to be more important for the classification because the statistically most significant parameters were selected in each step.

Important features

The common PCs are mainly related to asymmetry and structural properties rather than color (Table 4).

The linear classifier using only these five components achieved 0.933 AUC, 93.3% SE, and 91.1% SP using a leave-one-out cross-validation strategy. Dermatologists evaluate parallel patterns using the intensity distribution of the images, and they consider the peripheral area of the lesion as important.

We confirmed that our computer-based results, like those of the dermatologists, also focus on similar characteristics.

From Table 4, the linear classifier using only one PC could discriminate between melanomas and nevi with a high accuracy. Global texture properties used in the 2nd component




Figure 2. ROC curve for classification of acral volar melanomas.

Table 4. Interpretation of typical PCs and their diagnostic performance

PC	Interpretation	Numerical expression by top five constituting image features			
		$\lambda_i/\Sigma\lambda_i$ (%) ¹	SE (%) ²	SP (%) ²	AUC
2nd	$\cong 0.32$ diffX_T80+0.15 corr90_δ1/2-0.14 corr135_δ1/2.8-0.14 σ X_180-0.13 circ_T180 Asymmetry+global texture	12.8	93.3	65.1	0.833
9th	$\cong 0.21$ diffX_80+0.14 skewX_30-0.14 Brate_L1/40-0.14 Brate_L1/35-0.13 corr135_δ1/4 Peripheral+asymmetry+border	4.3	80.0	65.7	0.763
6th	$\cong 0.25$ skewS_tumor-norm+0.24 aveR_peri-norm+0.20 maxV_peri-0.20 maxR_tumor+0.15 σ S_tumor-peri Color (difference between tumor and skin)	7.0	66.7	70.4	0.711
1st	$\cong 0.23$ maxV_peri-0.19 corr90_δ1/4+0.19 maxR_peri+0.19 minR_peri+0.18 σ H_peri Peripheral+color	33.8	53.3	78.1	0.581
3rd	$\cong -0.18$ minH_peri+0.17 circ_T180-0.15 σ Y_T230+0.14 σ Y_T155 -0.13 ratio_T80 Asymmetry+peripheral	10.5	66.7	65.1	0.645

AUC, area under the receiver operating characteristic curve; PC, principal component; SE, sensitivity; SP, specificity.

¹Contribution ratio of the PC: ratio of eigenvalue λ_i to the total eigenvalues $\Sigma\lambda_i$.

²At which $SE \times SP$ is maximum value.

illustrate the periodicity of the structures present in the image. This meets our intuition, and it is considered that they contributed to discriminate effectively between ridges and furrows.

On the other hand, for the fibrillar pattern detector, most of the selected components were not significant; that is, they were low ranked (Table 2). Global texture features often used as a component of the parallel ridge detector were hardly selected by the fibrillar pattern detector, which seems not to extract the parallel patterns observed in ridges or furrows.

These results confirm that the features selected by acral volar classifiers showed a different trend. Our previous linear classifier, which was built on 1,258 non-acral dermoscopy images (1,050 melanocytic nevi and 198 melanomas) using the same method, achieved 0.914 AUC, 83.3% SE, and 85.8% SP using leave-one-out cross-validation. On the other hand, it achieved only 0.766 AUC, 70.0% SE, and 63.9% SP on the set of 199 acral volar images. We confirmed again from this experiment that discriminating acral volar melanomas by computer-based methods requires a specifically designed classifier that focuses more on asymmetry and textural properties of the image rather than only color.

Classification model

In this study, we used orthogonalized image features calculated by PCA. Because features extracted from the images are potentially correlated, PCA might reveal composite features that are more effective than their individual constituents.

In addition, it is well known that building a classifier with highly correlated parameters might result in poor classification accuracy for unseen data. On the other hand, we used linear models to classify acral melanomas and to identify specific dermoscopic structures. Although various sophisticated classifiers, such as artificial neural networks, logistic regression models, support vector machine classifiers, and so on, have been successfully used in the literature, using a linear model enables us to examine the relationship between input parameters and each class more clearly. From this point of view, the architecture of the classifier should be simple, taking into account the system generality, and we should focus more on finding effective input features rather than searching for the optimal classification model. The developed linear models achieved sufficiently high accuracy, and we consider that the extracted features were appropriate for classification.

Other issues

In this study, some images were from the same patients; however, as each image was from a different lesion and several evaluation methods provide almost equivalent results, we can ignore intrasubject correlation. On the other hand, our melanoma source was heavily concentrated in the Japanese population. Our classifier identified both Japanese and other cases accurately without significant difference, and therefore we can conclude that the diagnostic results did not depend on the source of images.

The presented system would classify the benign lesions with parallel ridge pattern such as Peutz-Jeghers syndrome as

malignant. However, we do not consider this to be a critical issue, as this system would be used primarily for screening.

On the other hand, we have to be careful not to miss melanomas that do not contain the parallel ridge pattern. Also, subcorneal hematoma and the like sometimes exhibit a similar appearance as the parallel ridge pattern. If clinical history is taken into account, the system could be made more accurate.

Our acral volar classifier successfully diagnosed 11 nevi and 5 melanomas that have no parallel pattern. We plan to investigate the issue of whether or not these lesions should be handled by an acral volar classifier. We will also continue collecting acral volar lesions to improve the generalization capability of the classifier.

The classifier described here will be available on our web-based screening system in the near future.

MATERIALS AND METHODS

Materials

Digital dermoscopy images of pigmented skin lesions (PSLs) were collected from four Japanese hospitals (Keio University Hospital, Tokyo; Toranomon Hospital, Tokyo; Shinshu University Hospital, Matsumoto; Inagi Hospital, Tokyo) and two European university hospitals (University of Naples, Italy; University of Graz, Austria) as part of the EDRA-CDROM (Argenziano *et al.*, 2000).

These were 24-bit color RGB images in JPEG (Joint Photographic Experts Group) format. A total of 213 acral volar dermoscopy images, 176 clinically equivocal nevi and 37 melanomas from 189 patients (75 male patients, 105 female patients, and 9 unknown), average and standard deviation of age of the patients 28.6 ± 22.2 years (max=87, min=0), were prepared. Note that the age information of 26 patients and 27 images was lost and these were not included in the above statistics. This experiment does not require intuitional approval or patient consent.

In this study, we used those images on which our tumor area extraction algorithm (Iyatomi *et al.*, 2006) was successful (169 benign nevi and 30 melanomas from 184 patients) (Table 1).

All cases were histopathologically or clinically diagnosed materials, and these results were used as the gold standard (training data) for building the classifier. We used all acral volar materials with established diagnosis stored in the above Japanese hospitals and CD resources (European cases).

Classification of acral lesions

We developed acral melanoma-nevus classifier and three detectors for typical patterns of acral volar lesions: parallel ridge pattern, parallel furrow pattern, and fibrillar pattern.

For melanoma-nevus classifier, the training signal of 1 or -1 was assigned to each melanoma and nevus case, respectively. Similarly, a training signal of 1 (positive) or -1 (negative) was assigned to each dermoscopic pattern.

The dermoscopic patterns were identified by three experienced dermatologists, and only those patterns on which at least two dermatologists agreed were considered.

Note that dermoscopic patterns were assessed independently of each other, and therefore some cases received multiple or no assignments.

We extracted tumor area with our tumor extraction algorithm and rotated the tumor object to align its major axis with the Cartesian x axis.

Then, we calculated a total of 428 image-related objective features. They can be roughly categorized into color, symmetry, border, and texture properties.

As color-related features, a total of 140 parameters were calculated: minimum (min), average (ave), maximum (max), standard deviation (σ), and skewness (skew) values in the RGB and HSV color space, respectively (subtotal 30), for the whole tumor area (tumor), periphery of the tumor area (peri), differences between the tumor area and the surrounding skin (tumor–skin), and differences between peripheral and surrounding skin (peri–skin).

In addition, a total of 20 color-related features were calculated: the number of colors (#color) in the tumor area and peripheral tumor area in the RGB and HSV color spaces quantized to 8^3 (= 512) and 16^3 (= 4,096) colors, respectively (subtotal 8), the average color of surrounding skin (R, G, B, H, S, V: subtotal 6), and average color differences between the peripheral tumor area and the inside of the tumor area (R, G, B, H, S, V: subtotal 6). Note that the peripheral part of the tumor is defined as the region inside the border that has an area equal to 30% of the tumor area.

In the symmetry category, a total of 80 features were calculated. We designed 10 intensity threshold values (T) from 5 to 230 with a step size of 25. In the extracted tumor area, thresholding was performed and the areas whose intensity was lower than the threshold were determined.

From each such area, we calculated eight features as follows: area ratio to original tumor size (ratio), circularity, differences of the center of gravity between original tumor (diffX, diffY), standard deviation of the distribution (σ_x , σ_y), and skewness of the distribution (skewX, skewY).

To quantify the border structure, a total of 32 features were calculated. The tumor areas are divided into eight equiangular regions. In each region, we defined a window whose size is $S \times S$ and center is on the border of the tumor. In each window, a ratio of color intensity between the inside and outside of the tumor and a gradient of color intensity were calculated in the blue and luminance channels (B_{ratio} , V_{ratio} , B_{grad} , V_{grad}), respectively. These were averaged over eight equiangular regions. We calculated four features for each of the eight window sizes (L): 1/5, 1/10, 1/15, 1/20, 1/25, 1/30, 1/35, and 1/40 of the length of the major axis of the dermoscopy image.

As for the texture features, a total of 176 parameters were calculated. We prepared 11 different-sized co-occurrence matrices with distance value δ ranging from 1/2 to 1/64 of the length of the major axis of the dermoscopy image. On the basis of each co-occurrence, the matrix, energy, moment, entropy, and correlation (corr) were calculated in four directions (0, 45, 90, and 135 degrees).

These 428 image features were transformed into $N(0,1)$ (zero mean and unit variance) and then orthogonalized using PCA.

Feature selection

From these orthogonalized features (=PCs), the parameters used in each classifier and detector were selected by an incremental stepwise method with a hypothesis test of Wilk's lambda (Everitt and Dunn, 1991). This method searches appropriate input parameters one after the other according to the statistical rule. In each step, a statistical F-test is performed and the parameter with the highest partial correlation efficient under $P < 0.05$ is selected. During this selection step, inefficient (statistically negligible;

$P > 0.10$) parameters are rejected from already selected parameters, if they exist. For each classifier and detector, this process was repeated until no further input satisfied the above-mentioned criteria.

Evaluation criteria

For melanoma–nevus classifier and three pattern detectors, the following evaluation criteria were used.

First, we calculated the standard deviation of mean estimated error (E) and a determination coefficient adjusted by degree of freedom (R^2) between the training signal and the developed linear model to evaluate the fitness of the developed classifier. Next, we evaluated the performance of the built classifiers using leave-one-out cross-validation, 10-fold cross-validation, and half-and-half tests, which split all data into training and evaluation in halves, and the performance was evaluated by averaging both combinations. Note that feature selection was performed using only the training set in each combination.

We used SE and SP measures and also plotted the ROC curve. In addition, positive predictive value and negative predictive value measures were also calculated. On the basis of the ROC curve, AUC was calculated. The AUC ranges from 0 to 1 and the greater its value the higher is the classification accuracy.

Analysis and evaluation tools

The PCA and evaluation of linear classifiers were performed using Matlab 13 with statistical toolbox software (MathWorks Inc., Natick, MA). The incremental stepwise input selection was performed using SPSS 14 software (SPSS Inc., Chicago, IL).

CONFLICT OF INTEREST

The authors state no conflict of interest.

ACKNOWLEDGMENTS

We appreciate Dr Toru Wakahara at Hosei University, who advises us on image processing. This research was supported in part by the Ministry of Education, Science, Sports and Culture, Grant-in-Aid for Young Scientists (B), 17790788, 2005–2006.

REFERENCES

- Argenziano G, Soyer HP, Chimenti S, Talamini R, Corona R, Sera F *et al.* (2003) Dermoscopy of pigmented skin lesions: results of a consensus meeting via the internet. *J Am Acad Dermatol* 48:679–93
- Argenziano G, Soyer HP, De Giorgi V, Piccolo D, Carli P, Delfino M *et al.* (2000) *Interactive atlas of dermoscopy CD*. Milan: EDRA Medical Publishing and New Media
- Blum A, Luedtke H, Ellwanger U, Schwabe R, Rassner G, Garbe C (2004) Digital image analysis for diagnosis of cutaneous melanoma. Development of a highly effective computer algorithm based on analysis of 837 melanocytic lesions. *Br J Dermatol* 151:1029–38
- Burroni M, Sbrano P, Cevenini G, Risulo M, Dell'Eva G, Barbini P *et al.* (2005) Dysplastic naevus vs. *in situ* melanoma: digital dermoscopy analyses. *Br J Dermatol* 152:679–84
- Celebi ME, Kingravi HA, Uddin B, Iyatomi H, Aslandogan YA, Stoecker WV *et al.* (2007) A methodological approach to the classification of dermoscopy images. *Comput Med Imaging Graph* 31:362–73
- Elbaum M, Kopf AW, Rabinovitz HS, Langley RG, Kamino H, Mihm MC Jr *et al.* (2001) Automatic differentiation of melanoma from melanocytic nevi with multispectral digital dermoscopy: a feasibility study. *J Am Acad Dermatol* 44:207–18

- Ercal F, Chawla A, Stoecker WV, Lee HC, Moss RH (1994) Neural network diagnosis of malignant melanoma from color images. *IEEE Trans Biomed Eng* 41:837-45
- Everitt BS, Dunn G (1991) *Applied multivariate data analysis*. London: Edward Arnold, 219-20
- Ganster H, Pinz A, Roehrer R, Wilding E, Binder M, Kitter H (2001) Automated melanoma recognition. *IEEE Trans Med Imaging* 20:233-9
- Hoffmann K, Gambichler T, Rick A, Kreutz M, Anschuetz M, Grunendick T et al. (2003) Diagnostic and neural analysis of skin cancer (DANAOS). A multicentre study for collection and computer-aided analysis of data from pigmented skin lesions using digital dermoscopy. *Br J Dermatol* 149:801-9
- Iyatomi H, Oka H, Masataka S, Miyake A, Kimoto M, Yamagami J et al. (2006) Quantitative assessment of tumour area extraction from dermoscopy images and evaluation of the computer-based methods for automatic melanoma diagnostic system. *Melanoma Res* 16:183-90
- Oka H, Hashimoto M, Iyatomi H, Tanaka M (2004) Internet-based program for automatic discrimination of dermoscopic images between melanoma and Clark nevi. *Br J Dermatol* 150:1041
- Rubegni P, Cevenini G, Burrioni M, Perotti R, Dell'Eva G, Sbrano P et al. (2002) Automated diagnosis of pigmented skin lesions. *Int J Cancer* 101:576-80
- Saida T, Miyazaki A, Oguchi S, Ishihara Y, Yamazaki Y, Murase S et al. (2004) Significance of dermoscopic patterns in detecting malignant melanoma on acral volar skin. *Arch Dermatol* 140:1233-8
- Saida T, Oguchi S, Miyazaki A (2002) Dermoscopy for acral pigmented skin lesions. *Clin Dermatol* 20:279-85
- Seidenari S, Pellacani G, Pepe P (1998) Digital videomicroscopy improves diagnostic accuracy for melanoma. *J Am Acad Dermatol* 39:175-81